



Inspection Report

LEGACY HEALTH
P. O. BOX 3950
PORTLAND, OR 97208

Customer ID: **1047**
Certificate: **92-R-0002**
Site: 001
LEGACY HEALTH - RES. ADMIN.

Type: ROUTINE INSPECTION
Date: 23-SEP-2021

2.33(b)(4) Critical

Attending veterinarian and adequate veterinary care.

An adverse event was reported to the IACUC. On 9/9/2021, a NHP was undergoing a procedure. During the procedure, a problem with respirations was noted. It was discovered that during setup system checks of the anesthesia machine, the popoff valve was left closed. The step of reopening the popoff valve post-pressure check was missed. By the time it was discovered, the NHP had suffered fatal barotrauma to the lungs.

Following approved procedural and safety protocols exactly is required for the safety of the animals and for the quality of the science.

This was corrected prior to the inspection by retraining all personnel responsible for administering anesthesia, and by ordering new safety valves for all the anesthesia machines that will be installed as soon as they arrive.

This inspection and exit interview were conducted with facility representatives.

Prepared By: GWYNN HALLBERG
USDA, APHIS, Animal Care
Title: VETERINARY MEDICAL
OFFICER

Date:
27-SEP-2021

Received by Title: IACUC Representative

Date:
27-SEP-2021



Species Inspected

Cust No	Cert No	Site	Site Name	Inspection
1047	92-R-0002	001	LEGACY HEALTH - RES. ADMIN.	23-SEP-2021

Count	Scientific Name	Common Name
000012	<i>Macaca mulatta</i>	RHESUS MACAQUE
000007	<i>Sus scrofa domestica</i>	DOMESTIC PIG / POTBELLY PIG / MICRO PIG
000031	<i>Ovis aries aries</i>	SHEEP INCLUDING ALL DOMESTIC BREEDS
000050	Total	